



NOTE Column change instructions are only to be used when maintenance is required and are not intended for practice purpose.



www.inogen.com



Insert column into the Inogen One[®] G₃ concentrator. Do not leave the column exposed; it should be inserted into the Inogen One[®] G₃ as soon as the dust caps have been removed.

Push the column into the device such that the column is fully seated into the Inogen One® G3 concentrator. The spring loaded latch button should fully return to the locked and closed position. Repeat steps 8-10 to install the other column.

CONTINUES ON BACK ->



Connect the AC power supply cord to the Inogen One[®] G₃ and plug the power supply AC cord into an electrical outlet. **Do Not Power ON the Inogen One[®] G₃ concentrator.**



Press and hold the light button for 10 seconds and screen will display "sieve reset". Release button once message is displayed on screen. Press the bell button once and screen will display "sieve reset successful". Press the power button to turn on the Inogen One® G₃, and use normally.

Your Equipment Provider:

Your Physician:



326 Bollay Drive, Goleta, CA 93117 1-877-4-INOGEN (Toll Free) 1-877-446-6436 (Toll Free) 1-805-562-0515 (International) info@inogen.net www.inogen.com

I N O G E N O N E^{® G3} Column Change Procedure





96-03989-00-01B

Column Change Guide